

Medical Infrared Thermometer JXB-178

Specifications:

- 1. Measuring Range: Human Body / Forehead: 32.0 ° C ~ 42.9 ° C (90 ° F ~ 109 ° F)
- 2. Accuracy: ± 0.3 ° C (0.54 ° F)
- 3. Measuring Distance: 5 ~ 8cm (2 ~ 5.9in)
- 4. Automatic Stop: 5sec.
- 5. Resolution: 0.1 ° C (0.1 ° F)
- 6. Response Time: 0.5sec
- 7. Operating Temperature: 10 ° C ~ 40 ° C (50 ° F ~ 140 ° F)
- 8. Relative Humidity Rate: less than or equal to 85%