



1. Humidity : 0 % ~ 100 % RH

2. Temperature : -20 °C ~ 80 °C (-4 °F ~ 176 °F)

3. Dew point : (-20 °C ~ 80 °C / -4 °F ~ 176 °F)

4. Wet bulb : (0 °C ~ 80 °C / 32 °F ~ 176 °F)

5. Humidity error :

+ 3 % RH (25 °C @ 20 % ~ 80 % RH).

+ 2.5 % RH (0 °C ~ 80 °C)

6. Temperature error :

+ 0.5 °C / + 0.9 °F (25 °C @ 20 % RH)

+ 0.8 °C / + 1.5 °F (0 °C ~ 80 °C)

7. Humidity resolution : 0.01 % RH, 0.01 °C / 0.01 °F.

8. Humidity stability : % RH : 10S (90 % +25 °C @ 20 % RH)